



3M™ Attest™ Mini Auto-reader 490M System* Quick Reference Guide

VH2O2 BI Preparation and Use

- Inspect**
Remove BI from pouch and reseal. Verify media ampoule is intact and indicator stripes on cap are blue.
- Pouch**
Place BI and chemical indicator (CI) in peel pouch indicated for use in vaporized hydrogen peroxide (VH2O2) sterilization processes.
- Process**
Seal pouch and place in sterilizer with white side facing up. Process load as recommended by sterilizer manufacturer.
- Identify**
After completion of cycle, inspect and verify media ampoule is intact and indicator stripes on cap have turned towards pink. Identify the BI by writing on the label.

STEAM and VH2O2 BI Activation and Incubation

- Crush**
Squeeze BI in activator to close cap and crush media ampoule.
- Flick**
Remove BI and flick wrist to ensure media flows to growth chamber.
- Inspect**
Verify media has transferred to BI growth chamber at the vial bottom.
- Incubate**
Place activated BI in any well. Display shows remaining minutes of incubation or test results. **DO NOT REMOVE BI until the (+) or (-) symbol indicates the test is complete.**

3M™ Attest™ Mini Auto-reader 490M

Positive Control BI
For both Steam and VH2O2, each day a processed BI is incubated, activate and incubate an unprocessed BI as a positive control. The control BI should be from the same lot code as the processed BI.

Results After Incubation
Control BI: Positive (+) **Acceptable**
Processed BI: Negative (-) **Acceptable Sterilization**
Processed BI: Positive (+) **Failed Sterilization**

Remaining Incubation Minutes

Touch Button (alarm off)

NOTES:

- Activate and incubate 1295 BI within 1 hour of completion of the VH2O2 sterilization cycle.
- Gloves are recommended when activating 1295 BI.
- Allow 1491 and 1492V BIs to cool 10 minutes prior to activation.

Interpretation of Solventum Sterilization Assurance Core Products

STEAM Indicator	Unexposed	Acceptable Exposed	VH2O2 Indicator	Unexposed	Acceptable Exposed Color Range
3M™ Attest™ BI 1491 and 1492V			3M™ Attest™ BI 1295		
3M™ Comply™ Indicator Tape 1322			3M™ Comply™ Indicator Tape 1228		
3M™ Attest™ Indicator Tape 1355			3M™ Attest™ Tri-Metric Chemical Indicator 1348/1348E		

STEAM Indicator	Unprocessed	Acceptable Processed PASS	Unacceptable Processed FAIL
3M™ Attest™ Steam Chemical Integrator 1243			
3M™ Comply™ Bowie-Dick Plus Test Pack 00135LF			
3M™ Attest™ eBowie-Dick Test System			

STEAM Indicator	Unprocessed	Processed	Processed PASS	Processed FAIL
3M™ Attest™ eBowie-Dick Test System				